

Table of contents

\USER

INVESTIGATORS

Anastasia

BANDA.2018.04.06

localizer_64ch
AAHead_Scout
HCP_MGH_T1w_MPR_vNav_setter
HCP_MGH_T2w_SPC_vNav_setter
HCP_MGH_T1w_MPR_vNav
HCP_CMRR_dMRI_dir98_AP
HCP_CMRR_dMRI_dir98_PA
HCP_CMRR_SpinEchoFieldMap_AP
HCP_CMRR_SpinEchoFieldMap_PA
HCP_CMRR_fMRI_rest_AP
HCP_CMRR_fMRI_rest_PA
HCP_CMRR_dMRI_dir99_AP
HCP_CMRR_dMRI_dir99_PA
HCP_CMRR_SpinEchoFieldMap_AP
HCP_CMRR_SpinEchoFieldMap_PA
HCP_CMRR_fMRI_rest_AP
HCP_CMRR_fMRI_rest_PA
HCP_CMRR_fMRI_gambling_AP
HCP_CMRR_fMRI_gambling_PA
HCP_CMRR_SpinEchoFieldMap_AP
HCP_CMRR_SpinEchoFieldMap_PA
HCP_CMRR_fMRI_faceMatching_AP
HCP_CMRR_fMRI_faceMatching_PA
HCP_CMRR_fMRI_conflict_AP
HCP_CMRR_fMRI_conflict_PA
HCP_CMRR_fMRI_conflict_AP
HCP_CMRR_fMRI_conflict_PA
HCP_MGH_T2w_SPC_vNav

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\localizer_64ch

TA: 0:12 PM: REF Voxel size: 0.5×0.5×7.0 mmPAT: Off Rel. SNR: 1.00 : fl

Properties

Prio recon	On
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Phase oversampling	0 %
FoV read	250 mm
FoV phase	100.0 %
Slice thickness	7.0 mm
TR	7.5 ms
TE	3.69 ms
Averages	2
Concatenations	3
Filter	Prescan Normalize, Elliptical filter
Coil elements	HC1-7

Contrast - Common

TR	7.5 ms
TE	3.69 ms
TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	20 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	2
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Contrast - Dynamic

Multiple series	Each measurement
-----------------	------------------

Resolution - Common

FoV read	250 mm
FoV phase	100.0 %
Slice thickness	7.0 mm
Base resolution	256
Phase resolution	91 %
Phase partial Fourier	Off
Interpolation	On

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Coronal
Phase enc. dir.	R >> L
FoV read	250 mm
FoV phase	100.0 %
Slice thickness	7.0 mm
TR	7.5 ms
Multi-slice mode	Sequential
Series	Interleaved
Concatenations	3

Geometry - AutoAlign

Slice group	1
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Position	L0.0 A20.0 H0.0 mm

Geometry - AutoAlign

Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Position	L0.0 A20.0 H0.0 mm
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Initial Position	L0.0 A20.0 H0.0
L	0.0 mm
A	20.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slice-sel.

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	7.5 ms
Concatenations	3
Segments	1

Physio - Cardiac

Tagging	None
Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	250 mm
FoV phase	100.0 %
Phase resolution	91 %

Physio - PACE

Resp. control	Off
Concatenations	3

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	20 deg
Measurements	1
Contrasts	1
TR	7.5 ms
TE	3.69 ms

Sequence - Part 1

Introduction	On
--------------	----

Sequence - Part 1

Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Allowed
Contrasts	1
Flow comp.	No
Multi-slice mode	Sequential
Bandwidth	320 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	None
RF pulse type	Fast
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
Allowed delay	0 s

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\AAHead_Scout

TA: 0:14 PM: REF Voxel size: 1.6×1.6×1.6 mmPAT: 3 Rel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L0.0 A10.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.15 ms
TE	1.37 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	3.15 ms
TE	1.37 ms
Flip angle	8 deg

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
Base resolution	160
Phase resolution	100 %
Slice resolution	69 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
Trajectory	Cartesian

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	24
Accel. factor 3D	1

Resolution - iPAT

Reference scan mode	Integrated
---------------------	------------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L0.0 A10.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.15 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	L0.0 A10.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off

System - Miscellaneous

Coil Select Mode	Off - AutoCoilSelect
------------------	----------------------

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Flip angle	8 deg
Measurements	1
Time to center	6.2 s

Inline - Inline

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	8 deg
Measurements	1

Inline - MapIt

Contrasts	1
TR	3.15 ms
TE	1.37 ms

Sequence - Part 1

Introduction	On
Dimension	3D
Asymmetric echo	Weak
Contrasts	1
Multi-slice mode	Sequential
Bandwidth	540 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Normal
Excitation	Non-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
------	-----

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_MGH_T1w_MPR_vNav_setter

TA: 0.2 s PM: FIX Voxel size: 8.0×8.0×8.0 mmRel. SNR: 1.00 : ABCD

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	L0.0 A25.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	32
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	8.00 mm
TR	9.9 ms
TE	4.6 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	BC

Contrast - Common

TR	9.9 ms
TE	4.6 ms
MTC	Off
Flip angle	2 deg
Fat suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	8.00 mm
Base resolution	32
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	6/8
Interpolation	Off

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	L0.0 A25.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	32
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	8.00 mm
TR	9.9 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	L0.0 A25.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A25.0 H0.0
L	0.0 mm
A	25.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---

System - Miscellaneous

Coil Select Mode	Default
------------------	---------

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	Isocenter
! Orientation	Transversal
! Rotation	0.00 deg
! A >> P	263 mm
! R >> L	350 mm
! F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	2.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	9.9 ms
Concatenations	1

Sequence - Part 1

Introduction	Off
Dimension	3D
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.26 ms
Bandwidth	5208 Hz/Px

Sequence - Part 2

EPI factor	32
RF pulse type	Normal
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On

Sequence - Special

Protocol filename	MPRAGE
-------------------	--------

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_MGH_T2w_SPC_vNav_setter

TA: 0.3 s PM: FIX Voxel size: 8.0×8.0×8.0 mmRel. SNR: 1.00 : ABCD

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	L0.0 A25.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	32
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	8.00 mm
TR	13.0 ms
TE	6.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	BC

Contrast - Common

TR	13.0 ms
TE	6.0 ms
MTC	Off
Flip angle	2 deg
Fat suppr.	Water excit. normal

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	8.00 mm
Base resolution	32
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	6/8
Interpolation	Off

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	L0.0 A25.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	32
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	8.00 mm
TR	13.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	L0.0 A25.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A25.0 H0.0
L	0.0 mm
A	25.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	Water excit. normal
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---

System - Miscellaneous

Coil Select Mode	Off - All
------------------	-----------

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	50.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	13.0 ms
Concatenations	1

Sequence - Part 1

Introduction	Off
Dimension	3D
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.26 ms
Bandwidth	5208 Hz/Px

Sequence - Part 2

EPI factor	32
RF pulse type	Normal
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On

Sequence - Special

Protocol filename	T2-SPACE
-------------------	----------

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_MGH_T1w_MPR_vNav

TA: 7:50 PM: REF Voxel size: 0.8×0.8×0.8 mmPAT: 2 Rel. SNR: 1.00 : tfl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
Slice oversampling	23.1 %
Slices per slab	208
FoV read	256 mm
FoV phase	93.8 %
Slice thickness	0.80 mm
TR	2400.0 ms
TE	2.18 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	2400.0 ms
TE	2.18 ms
Magn. preparation	Non-sel. IR
T1	1040 ms
Flip angle	8.0 deg
Fat suppr.	Water excit. fast
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	93.8 %
Slice thickness	0.80 mm
Base resolution	320
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	32
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	On
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	23.1 %
Slices per slab	208
FoV read	256 mm
FoV phase	93.8 %
Slice thickness	0.80 mm
TR	2400.0 ms
Multi-slice mode	Single shot
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L

System - Miscellaneous

Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Sagittal
Rotation	0.00 deg
A >> P	240 mm
F >> H	256 mm
R >> L	167 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	5.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	2400.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. IR
T1	1040 ms
Fat suppr.	Water excit. fast
Dark blood	Off
FoV read	256 mm
FoV phase	93.8 %
Phase resolution	100 %

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off

Inline - MIP

Save original images	On
----------------------	----

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Inline - Maplt

Save original images	On
Maplt	None
Flip angle	8.0 deg
Measurements	1
Contrasts	1
TR	2400.0 ms
TE	2.18 ms

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Allowed
Contrasts	1
Flow comp.	No
Multi-slice mode	Single shot
Echo spacing	7.9 ms
Bandwidth	220 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On
Turbo factor	300

Sequence - Special

Readout polarity	Positive
Nav. location	After
Apply moco to	parent and nav
Remeasure	24 TRs
Reacq. threshold	0.50
Feedback delay	80 ms
Moco ref. image	Use Temp Ref
K-space streaming	None
ABCD navigator	On
Add. grad time	0.0 ms
Apply freq to	parent and nav
Averaging	None

Sequence - Assistant

Mode	Off
------	-----

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_dMRI_dir98_AP

TA: 5:37 PM: REF Voxel size: 1.5×1.5×1.5 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	92
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
TR	3230 ms
TE	89.20 ms
Multi-band accel. factor	4
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	3230 ms
TE	89.20 ms
MTC	Off
Magn. preparation	None
Flip angle	78 deg
Refocus flip angle	160 deg
Fat suppr.	None
Grad. rev. fat suppr.	Enabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
Base resolution	140
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	92
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
TR	3230 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	None
Grad. rev. fat suppr.	Enabled
Special sat.	None

Geometry - Navigator**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	210 mm
R >> L	210 mm
F >> H	138 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3230 ms
Multi-band accel. factor	4

Physio - PACE

Resp. control	Off
Multi-band accel. factor	4

Diff - Neuro

Diffusion mode	Free
Diff. directions	98
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	3000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	On
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	Free
Diff. directions	98

Diff - Body

Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	3000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	40

Diff - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	On
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.69 ms
Bandwidth	1700 Hz/Px

Sequence - Part 2

EPI factor	140
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3840 us
Refocus pulse duration	7680 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Time-shifted MB RF	Off
SENSE1 coil combine	On
Invert RO/PE polarity	Off
Save reduced raw data	Off
PF omits higher k-space	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_dMRI_dir98_PA

TA: 5:37 PM: FIX Voxel size: 1.5×1.5×1.5 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	92
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
TR	3230 ms
TE	89.20 ms
Multi-band accel. factor	4
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	3230 ms
TE	89.20 ms
MTC	Off
Magn. preparation	None
Flip angle	78 deg
Refocus flip angle	160 deg
Fat suppr.	None
Grad. rev. fat suppr.	Enabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
Base resolution	140
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	92
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
TR	3230 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	None
Grad. rev. fat suppr.	Enabled
Special sat.	None

Geometry - Navigator**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	210 mm
R >> L	210 mm
F >> H	138 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3230 ms
Multi-band accel. factor	4

Physio - PACE

Resp. control	Off
Multi-band accel. factor	4

Diff - Neuro

Diffusion mode	Free
Diff. directions	98
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	3000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	On
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	Free
Diff. directions	98

Diff - Body

Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	3000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	40

Diff - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	On
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.69 ms
Bandwidth	1700 Hz/Px

Sequence - Part 2

EPI factor	140
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3840 us
Refocus pulse duration	7680 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Time-shifted MB RF	Off
SENSE1 coil combine	On
Invert RO/PE polarity	On
Save reduced raw data	Off
PF omits higher k-space	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_SpinEchoFieldMap_AP

TA: 0:32 PM: FIX Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
TE	66.00 ms
Multi-band accel. factor	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	8000 ms
TE	66.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain

System - Miscellaneous

Coil Select Mode	Off - All
------------------	-----------

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	8000 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	Off
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - Special

SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_SpinEchoFieldMap_PA

TA: 0:32 PM: FIX Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
TE	66.00 ms
Multi-band accel. factor	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	8000 ms
TE	66.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain

System - Miscellaneous

Coil Select Mode	Off - All
------------------	-----------

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	8000 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	Off
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - Special

SENSE1 coil combine	Off
Invert RO/PE polarity	On
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_fMRI_rest_AP

TA: 5:46 PM: FIX Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	420
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	420
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
------------	-----

Sequence - Part 2

Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_fMRI_rest_PA

TA: 5:46 PM: FIX Voxel size: 2.0x2.0x2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	420
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	420
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
------------	-----

Sequence - Part 2

Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_dMRI_dir99_AP

TA: 5:41 PM: FIX Voxel size: 1.5×1.5×1.5 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	92
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
TR	3230 ms
TE	89.20 ms
Multi-band accel. factor	4
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	3230 ms
TE	89.20 ms
MTC	Off
Magn. preparation	None
Flip angle	78 deg
Refocus flip angle	160 deg
Fat suppr.	None
Grad. rev. fat suppr.	Enabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
Base resolution	140
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	92
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
TR	3230 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	None
Grad. rev. fat suppr.	Enabled
Special sat.	None

Geometry - Navigator**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	210 mm
R >> L	210 mm
F >> H	138 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3230 ms
Multi-band accel. factor	4

Physio - PACE

Resp. control	Off
Multi-band accel. factor	4

Diff - Neuro

Diffusion mode	Free
Diff. directions	99
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	3000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	On
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	Free
Diff. directions	99

Diff - Body

Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	3000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	40

Diff - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	On
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.69 ms
Bandwidth	1700 Hz/Px

Sequence - Part 2

EPI factor	140
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3840 us
Refocus pulse duration	7680 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Time-shifted MB RF	Off
SENSE1 coil combine	On
Invert RO/PE polarity	Off
Save reduced raw data	Off
PF omits higher k-space	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_dMRI_dir99_PA

TA: 5:41 PM: FIX Voxel size: 1.5×1.5×1.5 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	92
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
TR	3230 ms
TE	89.20 ms
Multi-band accel. factor	4
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	3230 ms
TE	89.20 ms
MTC	Off
Magn. preparation	None
Flip angle	78 deg
Refocus flip angle	160 deg
Fat suppr.	None
Grad. rev. fat suppr.	Enabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
Base resolution	140
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	92
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
TR	3230 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	None
Grad. rev. fat suppr.	Enabled
Special sat.	None

Geometry - Navigator**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	210 mm
R >> L	210 mm
F >> H	138 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3230 ms
Multi-band accel. factor	4

Physio - PACE

Resp. control	Off
Multi-band accel. factor	4

Diff - Neuro

Diffusion mode	Free
Diff. directions	99
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	3000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	On
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	Free
Diff. directions	99

Diff - Body

Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	3000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	40

Diff - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	On
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.69 ms
Bandwidth	1700 Hz/Px

Sequence - Part 2

EPI factor	140
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3840 us
Refocus pulse duration	7680 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Time-shifted MB RF	Off
SENSE1 coil combine	On
Invert RO/PE polarity	On
Save reduced raw data	Off
PF omits higher k-space	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_SpinEchoFieldMap_AP

TA: 0:32 PM: FIX Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
TE	66.00 ms
Multi-band accel. factor	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	8000 ms
TE	66.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain

System - Miscellaneous

Coil Select Mode	Off - All
------------------	-----------

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	8000 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	Off
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - Special

SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_SpinEchoFieldMap_PA

TA: 0:32 PM: FIX Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
TE	66.00 ms
Multi-band accel. factor	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	8000 ms
TE	66.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain

System - Miscellaneous

Coil Select Mode	Off - All
------------------	-----------

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	8000 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	Off
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - Special

SENSE1 coil combine	Off
Invert RO/PE polarity	On
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_fMRI_rest_AP

TA: 5:46 PM: FIX Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	420
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	420
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
------------	-----

Sequence - Part 2

Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_fMRI_rest_PA

TA: 5:46 PM: FIX Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	420
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	420
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
------------	-----

Sequence - Part 2

Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_fMRI_gambling_AP

TA: 3:02 PM: FIX Voxel size: 2.0x2.0x2.0 mmPAT: Off Rel. SNR: 1.00 : efpid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	215
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	215
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
------------	-----

Sequence - Part 2

Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_fMRI_gambling_PA

TA: 3:02 PM: FIX Voxel size: 2.0x2.0x2.0 mmPAT: Off Rel. SNR: 1.00 : efpid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	215
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	215
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
------------	-----

Sequence - Part 2

Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_SpinEchoFieldMap_AP

TA: 0:32 PM: FIX Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
TE	66.00 ms
Multi-band accel. factor	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	8000 ms
TE	66.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain

System - Miscellaneous

Coil Select Mode	Off - All
------------------	-----------

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	8000 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	Off
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - Special

SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_SpinEchoFieldMap_PA

TA: 0:32 PM: FIX Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
TE	66.00 ms
Multi-band accel. factor	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	8000 ms
TE	66.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain

System - Miscellaneous

Coil Select Mode	Off - All
------------------	-----------

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	8000 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	Off
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	3
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - Special

SENSE1 coil combine	Off
Invert RO/PE polarity	On
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_fmRI_faceMatching_AP

TA: 5:34 PM: FIX Voxel size: 2.0x2.0x2.0 mmPAT: Off Rel. SNR: 1.00 : efpid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	405
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	405
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
------------	-----

Sequence - Part 2

Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_fmri_faceMatching_PA

TA: 5:34 PM: FIX Voxel size: 2.0x2.0x2.0 mmPAT: Off Rel. SNR: 1.00 : efpid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	405
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	405
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
------------	-----

Sequence - Part 2

Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_fmri_conflict_AP

TA: 3:54 PM: FIX Voxel size: 2.0x2.0x2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	280
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	280
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
------------	-----

Sequence - Part 2

Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_fmri_conflict_PA

TA: 3:54 PM: FIX Voxel size: 2.0x2.0x2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	280
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	280
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
------------	-----

Sequence - Part 2

Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_fmri_conflict_AP

TA: 3:54 PM: FIX Voxel size: 2.0x2.0x2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	280
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	280
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
------------	-----

Sequence - Part 2

Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_CMRR_fmri_conflict_PA

TA: 3:54 PM: FIX Voxel size: 2.0x2.0x2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	280
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P3.0 H6.0
R	0.0 mm
P	3.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-20.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P3.0 H6.0 mm
Orientation	T > C-20.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	280
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
------------	-----

Sequence - Part 2

Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Save reduced raw data	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\INVESTIGATORS\Anastasia\BANDA.2018.04.06\HCP_MGH_T2w_SPC_vNav

TA: 6:58 PM: FIX Voxel size: 0.8x0.8x0.8 mmPAT: 2 Rel. SNR: 1.00 : spc

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	208
FoV read	256 mm
FoV phase	93.8 %
Slice thickness	0.80 mm
TR	3200 ms
TE	564 ms
Averages	1.0
Concatenations	1
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	3200 ms
TE	564 ms
MTC	Off
Magn. preparation	None
Fat suppr.	None
Blood suppr.	Off
Restore magn.	Off

Contrast - Dynamic

Averages	1.0
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	93.8 %
Slice thickness	0.80 mm
Base resolution	320
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Allowed
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	32
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	On
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	208
FoV read	256 mm
FoV phase	93.8 %
Slice thickness	0.80 mm
TR	3200 ms
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm

System - Miscellaneous

MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Sagittal
Rotation	0.00 deg
A >> P	240 mm
F >> H	256 mm
R >> L	167 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.181081 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	5.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
Trigger delay	0 ms
TR	3200 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	256 mm
FoV phase	93.8 %
Phase resolution	100 %

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off

Inline - MIP

MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Flow comp.	No
Echo spacing	3.86 ms
Adiabatic-mode	Off
Bandwidth	744 Hz/Px

Sequence - Part 2

Echo train duration	1166 ms
RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
Flip angle mode	T2 var
Turbo factor	314

Sequence - Special

Include nav.	On
Apply moco to	parent and nav
Remeasure	18 TRs
Reacq. threshold	0.50
Feedback delay	80 ms
Moco ref. image	Use Temp Ref
K-space streaming	None
ABCD navigator	On
Apply freq to	parent and nav

Sequence - Assistant

Allowed delay	0 s
---------------	-----